



Revision No.: ZJ0007
Issue Date: Jul 1st, 2012

Polyomavirus BK (PBK) Real Time PCR Kit User Manual

For In Vitro Diagnostic Use Only

REF OD-0102-01

For use with LightCycler1.0/2.0 Instrument

EC **REF**

Obelis S.A.
Boulevard Général Wahis 53
1030 Brussels, BELGIUM
Tel: +(32) 2.732.59.54
Fax: +(32) 2.732.60.03
E-Mail : mail@obelis.net



Shanghai ZJ Bio-Tech Co., Ltd.
www.liferiver.com.cn Tel: +86-21-34680596
trade@liferiver.com.cn Fax: +86-21-34680595
2nd floor, No.15 Building, No.188 Xijunhuan road,
Pujiang Hi-tech Park Shanghai China



1. Intended Use

Polyomavirus BK (PBK) real time PCR kit is used for the detection of Polyomavirus BK virus in serum, plasma or urine sample by using real time PCR systems.

2. Principle of Real-Time PCR

The principle of the real-time detection is based on the fluorogenic 5' nuclease assay. During the PCR reaction, the DNA polymerase cleaves the probe at the 5' end and separates the reporter dye from the quencher dye only when the probe hybridizes to the target DNA. This cleavage results in the fluorescent signal generated by the cleaved reporter dye, which is monitored real-time by the PCR detection system. The PCR cycle at which an increase in the fluorescence signal is detected initially (Ct) is proportional to the amount of the specific PCR product. Monitoring the fluorescence intensities during Real Time allows the detection of the accumulating product without having to re-open the reaction tube after the amplification.

3. Product Description

Polyomavirus BK virus (PBK) is discovered from the urine of a renal transplant recipient in 1971. The polyomavirus is a subfamily of the papovavirus family. PBK infections occur in early childhood. Primary infections with PBK are essentially harmless but the viruses tend to persist indefinitely in the infected individual. The virus remains latent in the kidney and in B lymphocytes after primary infection. PBK infection has been linked to occasional cases of cystitis in immunocompetent children, to glomerulonephritis in immunodeficient children, and to haemorrhagic cystitis in bone marrow transplant recipients. Polyomavirus BK real time PCR kit contains a specific ready-to-use system for the detection of Polyomavirus BK by polymerase chain reaction (PCR) in the real-time PCR system. The master contains reagents and enzymes for the specific amplification of Polyomavirus BK DNA. Fluorescence is emitted and measured by the real time systems' optical unit during PCR. The detection of amplified Polyomavirus BK DNA fragment is performed in fluorimeter **channel 530nm** with the fluorescent quencher BHQ1. DNA extraction buffer is available in the kit and serum samples are used for DNA extraction. In addition, the kit contains a system to identify possible PCR inhibition by measuring the 560nm fluorescence of the internal control (IC).

4. Kit Contents

Ref.	Type of Reagent	Presentation	25rxns
1	DNA Extraction Buffer	1 vial, 1.8ml	
2	PBK Reaction Mix	1 vial, 450µl	
3	PCR Enzyme Mix	1 vial, 12µl	
4	Molecular Grade Water	1 vial, 400µl	
5	Internal Control (IC)	1 vial, 30µl	
6	PBK Positive Control	1 vial, 30µl	

Analysis sensitivity: 1×10⁴ copies/ml

Note: Analysis sensitivity depends on the sample volume, elution volume, nucleic acid extraction methods and other factors. If you use the DNA extraction buffer in the kit, the analysis sensitivity is the same as it declares. However, when the sample volume is dozens or even hundreds of times greater than elution volume by some concentrating method, it can be much higher.

5. Storage

- All reagents should be stored at -20°C. Storage at +4°C is not recommended.
- All reagents can be used until the expiration date indicated on the kit label.
- Repeated thawing and freezing (> 3x) should be avoided, as this may reduce the sensitivity of the assay.
- Cool all reagents during the working steps.
- Reaction Mix should be stored in the dark.

6. Additionally Required Materials and Devices

- Biological cabinet
- Real time PCR system
- Desktop microcentrifuge for "eppendorf" type tubes (RCF max. 16,000 x g)
- Vortex mixer
- Real time PCR reaction tubes/plates
- Cryo-container
- Pipets (0.5 µl – 1000 µl)
- Sterile filter tips for micro pipets
- Sterile microtubes
- Disposable gloves, powderless
- Biohazard waste container
- Refrigerator and freezer
- Tube racks

7. Warnings and Precaution

- Carefully read this instruction before starting the procedure.
- For in vitro diagnostic use only.
 - This assay needs to be carried out by skilled personnel.
 - Clinical samples should be regarded as potentially infectious materials and should be prepared in a laminar flow hood.
 - This assay needs to be run according to Good Laboratory Practice.
 - Do not use the kit after its expiration date.
 - Avoid repeated thawing and freezing of the reagents, this may reduce the sensitivity of the test.
 - Once the reagents have been thawed, vortex and centrifuge briefly the tubes before use.
 - Prepare quickly the Reaction mix on ice or in the cooling block.
 - Set up two separate working areas: 1) Isolation of the RNA/ DNA and 2) Amplification/ detection of amplification products.
 - Pipets, vials and other working materials should not circulate among working units.
 - Use always sterile pipette tips with filters.
 - Wear separate coats and gloves in each area.
 - Do not pipette by mouth. Do not eat, drink, smoke in laboratory.
 - Avoid aerosols

8. Sample Collection, Storage and transport

- Collect samples in sterile tubes;
- Specimens can be extracted immediately or frozen at -20°C to -80°C.
- Transportation of clinical specimens must comply with local regulations for the transport of etiologic agents

9. Procedure

9.1 DNA-Extraction

DNA extraction buffer is supplied in the kit, please thaw the buffer thoroughly and spin down briefly in the centrifuge before use. It's better to use commercial kits for nucleic acid extraction.

9.1.1 Serum or plasma sample

- 1) Pipet 50µl serum or plasma to a new 0.5ml tube, add 50µl DNA extraction buffer, close the tube then vortex for 10 seconds. Spin down briefly in a table centrifuge.
- 2) Incubate the tube for 10 minutes at 100°C.
- 3) Centrifuge the tube at 13000rpm for 10 minutes. The supernatant contains the DNA extracted and can be used for PCR template.

9.1.2 Urine sample

- 1) Take 1.5 ml sample to a tube, Centrifuge the tube at 13000rpm for 2 minutes, carefully remove and discard supernatant from the tube without disturbing the pellet.
- 2) Add 1.0ml normal saline then vortex vigorously. Centrifuge the tube at 15000rpm for 5 minutes, carefully remove and discard supernatant from the tube without disturbing the pellet.
- 3) Add 50µl DNA extraction buffer, lose the tube then vortex for 10 seconds.. Spin down briefly in a table centrifuge.
- 4) Incubate the tube for 10 minutes at 100°C.
- 5) Centrifuge the tube at 13000rpm for 5 minutes. The supernatant contains the DNA extracted and can be used for PCR template.

Attention:

- A.** During the incubation, make sure the tube is not open. Since the vapor will volatilize into the air and may cause contamination if the sample is positive.
- B.** The extraction sample should be used in 3 hours or stored at -20°C for one month.
- C.** DNA extraction kits are available from various manufacturers. You may use your own extraction systems or the commercial kit based on the yield. For the DNA extraction, please comply with the manufacturer's instructions.


9.2 Internal Control

It is necessary to add internal control (IC) in the reaction mix. Internal Control (IC) allows the user to determine and control the possibility of PCR inhibition.

Add the internal control (IC) 1µl/rxn and the result will be shown in the 560nm.

9.3 PCR Protocol

The Master Mix volume for each reaction should be pipetted as follows:



※PCR system without 560nm channel may be treated with 1µl Molecular Grade Water instead of 1µl IC.

- 1) The volumes of Reaction Mix and Enzyme Mix per reaction multiply with the number of samples, which includes the number of controls, standards, and sample prepared. Molecular Grade Water is used as the negative control. For reasons of unprecise pipetting, always add an extra virtual sample. Mix completely then spin down briefly in a centrifuge.
- 2) Pipet **18µl** Master Mix with micropipets of sterile filter tips to each of the *Real time* PCR reaction plate/tubes. Separately add **2µl** DNA sample, positive and negative controls to different reaction plate/tubes. Immediately close the plate/tubes to avoid contamination.
- 3) Spin down briefly in order to collect the Master Mix in the bottom of the reaction tubes.
- 4) Perform the following protocol in the instrument:

37°C for 2min	1cycle
94°C for 2min	1cycle
93°C for 5sec, 60°C for 30sec (Fluorescence measured at 60°C)	40cycles

Selection of fluorescence channels	
530nm	Target Nucleic Acid
560nm	IC

10.Threshold setting: Choose **Arithmetic** as back ground and **none** as Noise Band method, then adjust the Noise band just above the maximum level of molecular grade water, and adjust the threshold just under the minimum of the positive control.

11.Quality control: Negative control, positive control, internal control and QS curve must be performed correctly, otherwise the sample results is invalid.

Channel	Crossing point value	
	530nm	560nm
Control	Blank	25~35
Molecular Grade Water	Blank	25~35
Positive Control(qualitative assay)	≤35	—

12. Data Analysis and Interpretation

The following results are possible:

	Crossing point value		Result Analysis
	530nm	560nm	
1#	Blank	25~35	Below the detection limit or negative
2#	≤35	—	Positive
3#	35~40	25~35	Re-test; If it is still 35~40, report as 1#
4#	Blank	Blank	PCR Inhibition; No diagnosis can be concluded.

For further questions or problems, please contact our technical support at trade@liferiver.com.cn